

# Certificate



## Safe and hygienic reprocessing of laundry with Miele Professional washing machines. from the Benchmark, 24/32 kg and barrier machine ranges.

The Miele disinfection programmes installed ex works on Benchmark (PW 413 MopStar, PW 418 MopStar, PW 811, PW 814 and PW 818), 24/32 kg (PW 6241 and PW 6321) and barrier (PW 6163, PW 6243 and PW 6323) machines comply fully with guidelines issued by the Robert Koch Institute (RKI) and the Association for Applied Hygiene (VAH).

These strict requirements refer, inter alia, to the maintenance of temperatures in the wash liquor. These machines guarantee that selected temperatures are maintained for the time specified in the relevant programmes. In disinfection programmes, in particular, temperatures are maintained with tightly controlled hysteresis above the selected value in order to prevent any drop in temperature below the target temperature.

These machines also include further safety functions to ensure proper disinfection or to inform users if disinfection has not been achieved:

- **Programme lock: The machine door can only be opened at the end of a wash cycle.**
- **If, in the event of a fault, the requisite temperature is not maintained, the appropriate programme block is recommenced from the beginning.**
- **Programme abort in the event of an empty supply container and 'Supply container empty' message**
- **No further water intake via detergent dispenser after disinfection phase**

Depending on the machine version, controls are freely programmable, i.e. clients are able to customise programmes to suit individual requirements. Modifications to disinfection programmes can have significant impact on the hygienic reprocessing of laundry. Hence all claims relating to proper hygiene and disinfection relate solely to genuine Miele wash programmes.

Alongside technical specifications, reliable hygiene also requires that the following is ensured by the operator or the machine's owner:

When chemo-thermal processes are used:

- **Observance of instructions and information contained in operating instructions**
- **Use of detergent suitable for disinfection**
- **Compliance with instructions issued by detergent manufacturers, in particular relating to load-to-liquor ratios, in order to ensure the correct amount of detergent is dispensed**

Chemo-thermal disinfection programmes are set to a load-to-liquor ratio of approx. 1:5 for a nominal load with average absorbency.

A load-to-liquor ratio of approx. 1:4 to 1:5 applies to thermal disinfection programmes.

Precise parameters in chemo-thermal disinfection programmes with exact temperature and dispensing points as well as a load-to-liquor ratio adapted to the load in hand can be set by a service technician during machine commissioning.

The 'Load-to-liquor ratio' programming mode can be selected with accuracy and automatically in combination with the weighing system as an optional extra and a flowmeter for each load size and laundry type.

When thermal processes are used:

- **Observance of instructions and information contained in operating instructions**

Please note the enclosed excerpts from the operating instructions of Miele machines.

- **Validity:** 3 years | Only for models from the Benchmark series (PW 413 MopStar, PW 418 MopStar, PW 811, PW 814 and PW 818), '24/32 kg' range (PW 6241 and PW 6321) and barrier machines (PW 6163, PW 6243 und PW 6323)
- **Authors:** M. Röhl and J. Sielmann
- **Source:** Miele & Cie. KG
- **Scope and term of validity:** 01.11.2019 until 01.11.2022

Tim van Zoest, Laboratory Manager, Miele Lehrte  
Miele & Cie. KG, Carl-Miele-Straße 29, D-33332 Gütersloh

Enclosure:

## Disinfection - general information

### What is disinfection?

Deactivation and reduction in number of pathogens transmitting diseases to such an extent that the transmission of diseases and infection is no longer to be anticipated.

### Thermal disinfection

Thermal disinfection is carried out by the action of a defined temperature over a holding time.

In the list, according to § 18 IFSG, this is 90°C over 10 mins. or 85°C over 15 mins. With a load-to-liquor ratio of 1 - 4 to 1 - 5 with the activity range AB (cf. Section on 'Disinfection processes according to activity range').

The thermal disinfection is independent of the detergent used, which does not have to be classified.

### Chemo-thermal disinfection

Chemo-thermal disinfection differs from thermal disinfection in terms of a lower liquor temperature during the disinfection phase.

Compensation for the low temperature is ensured through the use of listed washing and disinfecting components. It is important that these products are used exactly according to the listings.

The suds temperature, the load-to-liquor ratio and the time of dispensing must be strictly adhered to.

For disinfection wash programmes there are two lists in which the necessary process parameters with regard to washing and disinfecting agents are listed.

### RKI list (Robert Koch Institute)

Describes procedures for disinfection according to § 18 Infection Protection Act. Requirements

These processes are only required when a health authority prescribes the use of disinfection in the event of the outbreak of an epidemic relating to a notifiable, transferable disease where there is a risk of items being contaminated. The RKI list describes both thermal and chemo-thermal disinfection processes.

### Disinfection process according to areas of impact

Disinfection procedure differ by the relevant areas of impact A, B and C.

– Area of impact A:

Suitable for the destruction of vegetative bacteria, including mycobacteria, fungus and fungal spores.

– Area of impact B:

Suitable for deactivating viruses.

– Area of impact C:

To deactivate anthrax spores.

In the event of officially directed disinfection, the first wash liquor may not be discharged until after disinfection has been carried out.

Enclosure:

## VAH list

VAH - Association for Applied Hygiene  
(formerly DGHM = German Society for Hygiene and Microbiology)

The VAH/DGHM list describes the procedures required for routine prophylactic disinfection, and in particular for the prevention of infection in hospitals, doctors' offices, public areas and other areas where there may be the risk of infection spreading.

With respect to a notifiable disinfection, § 18 of the Infection Protection Act is invoked.

Special validated procedures and machines are to be used for surgical textiles as medical devices. The procedures and machines listed here are not medical devices.

### Avoidance of contamination or recontamination

- Disinfect your hands before using the machine.
- Use clean / disinfected transport containers.
- Dispense washing and disinfecting agents in accordance with instructions.
- Observe regulations regarding the wearing of protective clothing.
- Ensure that water is bacteria-free when using a supply from water treatment units.
- Clean detergent dispensing drawers, detergent compartments and siphons before use.

### Cleaning dispensing drawers and siphons

The detergent dispensing drawer is subject to constant moisture. It needs to be cleaned regularly to prevent the build-up of water-borne bacteria.

- Clean the detergent dispensing drawer, detergent compartments and siphon thoroughly with hot water as required to remove detergent residues and encrustation.

It is the operator's responsibility to ensure that the standard of disinfection in thermal as well as chemo-thermal procedures is maintained (in accordance with local and national health and safety regulations) by carrying out suitable validation tests. The procedures should be checked periodically and thermoelectrically by means of loggers or bacteriologically using bio-indicators. The operator must pay particular attention to observing the process parameter of temperature and, in the case of chemo-thermal programmes, the concentration. Disinfection programmes must not be interrupted, as this can have a limiting effect on the disinfection result.