

16.02.2023

Miele spray nozzle system for machine reprocessing of masks and demand valves

Confirmation that the Miele PG 8063 washer is suitable for cleaning and disinfecting the following Dräger masks and regulators.

Fullface masks: FPS 7000 RA, PE, P, ESA Panorama Nova RA, PE, P, ESA (without any coatings)
X-plore 5500/6000 Series; CDR 4500 Series

Halfmasks: X-Plore 2/3/4000 (not approved with Dr.Weigert Dekonta AF)

LDV: PSS LDV RD40, PSS LDVM45x3; PSS LDV ESA, PSS LDV Push-In

This document is to be seen as a supplement to the current instructions for use for the appropriate Dräger product which will be cleaned. The instructions for use are to be regarded as leading documents in which the parameters to be observed are described.

Instructions for use to be observed

- Machine reprocessing of full-face masks with spray process (9031089)
- Lung Demand Valve Machine cleaning and disinfecting (3363265)

To determine the suitability, a selection of Dräger products was reprocessed in the Miele spray nozzle system. After completion of the tests, the suitability for use of the masks and demand valves was confirmed.

The Miele spray nozzle system is thus approved in compliance with the parameters specified in the instructions for use for the respective product to be cleaned.

Cleaning system

- Washer PG 8063
- With the special baskets APFD 200 for 6 LDVs and APFD 201 for 6 respiratory masks

Chemistry and parameters

Diversey Suma® Jade Pur-Eco L8/ Suma® Med Neutral	Dr. Weigert neodisher® MediClean Forte/ neodisher® Polyklar	Dr. Weigert neodisher® Dekonta AF/ neodisher® Polyklar
Duration: 5 mins. / min. 25 secs. Concentration: 0.4 % / 0,05 % - 0.1 % Temperature: 50°C - 55 °C	Duration: 5 mins. to 22 mins. / min. 25 secs. Concentration: 0.5 % - 1.0 % / 0.05 % - 0.1 % Temperature: 50 °C - 55 °C	Duration: 5 mins. to 22 mins. / min. 25 secs. Concentration: 1.0 % / 0.05 % - 0.1 % Temperature: 50 °C - 55 °C



Contact information:

Simon Vogt
Global Product & Application Manager
Cleaning Solutions

Dräger Safety AG & Co. KGaA
Revalstraße 1
23560 Lübeck

Mobil +49 172 4000 934
Tel +49 451 882 1708

Simon.Vogt@draeger.com
www.draeger.com

Dräger Safety AG & Co. KGaA
Revalstraße 1
23560 Lübeck, Germany
Phone +49 451 882-0
Fax +49 451 882-2080
info@draeger.com
www.draeger.com
UID no. DE8128834
WEEE Reg. No. DE13636930

Bank details:
Commerzbank AG, Lübeck
IBAN DE29 2304 0022 0014 6803 00
BIC COBADEFF230
Deutsche Bank AG, Lübeck
IBAN DE75 2307 0710 0030 2109 00
BIC DEUTDEHH222

Registered office of the company: Lübeck
Commercial Register:
Lübeck Local Court HRB 4097 HL
General Partner:
Dräger Safety Verwaltungs AG
Registered office: Lübeck
Commercial Register:

Chairman of the Supervisory Board of Dräger Safety AG & Co. KGaA and Dräger Safety Verwaltungs AG: Stefan Lauer
Board of Directors:
Stefan Draeger (Chairman)
Rainer Klug
Gert-Hartwig Lescow
Dr. Reiner Piske
Anton Schrofner