

## ProCare

### Dent 10 A

## Detergent


### Application:

Liquid concentrate, alkaline. For the automated reprocessing of dental instruments in Miele Professional washer-disinfectors.

*ProCare Dent 10 A* supports the removal of organic soiling such as blood and saliva and prevents any redeposition. Coarse residues from composites, dental cement and dental impression compound should be removed before they have had time to set.

*ProCare Dent 10 A* is not suitable for cleaning transmission instruments. The use of the *ProCare Dent 10 MA* liquid detergent is recommended for this purpose.

### Quality features:

*ProCare Dent 10 A* complies with European guidelines for medical devices. 

If a serious incident occurs with the product, report it to the manufacturer and the relevant national authority.

### Product properties:


- Excellent material compatibility and corrosion protection for instruments and trays made from stainless steel
- Does not contain oxidants or surfactants

### Use and dispensing:

Miele *ProCare Dent 10 A* is formulated for use in processes in Miele Professional washer-disinfectors.

- Precise and economical dispensing via dispenser pump.
- Suitable dosing dispensers should be used.
- Application concentration: 3 ml/l
- Program selection: Vario TD

### Program sequence:



### vario TD process:

- 1 Pre-wash**  
with cold water to remove coarse soiling (e.g. blood)
- 2 Main wash**  
with *ProCare Dent 10 A* at 130 °F / 55 °C (dispensing temperature 105 °F / 40 °C)
- 3 Neutralization**  
with *ProCare Dent 30 P*
- 4 Intermediate rinse**  
with cold water
- 5 Thermal disinfection**  
at 200 °F / 93 °C with 5-minute holding time in the final rinse
- 6 Drying**

The product in solution is to be completely rinsed off using water. In Miele Professional washer-disinfectors, this is guaranteed by the Vario TD program when properly used (e.g. correct loading). Verification is provided by measuring the increase of the conductivity in the final rinse during commissioning and routine maintenance. Miele recommends the use of fully demineralized water in the final rinse in order to protect the load and the washer-disinfector against stains and deposits.

### Please note!

- Do not mix with other products.
- Contact Miele Service when switching products.
- Reprocessing should comply with all ordinances pursuant to medical device regulations and should be performed with appropriate validated processes.
- Observe recommendations on reprocessing issued by medical device manufacturers with respect to the provisions of DIN EN ISO 17664.
- Operating instructions on Miele Professional washer-disinfectors must be observed.
- For commercial applications only.

## ProCare

### Dent 10 A

## Detergent


#### Technical data:

- Density (at 68 °F / 20 °C): approx. 1.5 g/cm<sup>3</sup>
- pH-value (determined using fully demineralized water, 68 °F / 20 °C) 3 ml/l: approx. 11.6

#### Ingredients according to Regulation (EC) 648/2004 on detergents:

- 15 % to 30 % phosphates

#### Storage:

- Storage temperature: 5 °F to 86 °F / -15 °C to 30 °C
- Can be used for 3 years under proper storage conditions
- Use-by date: See label next to hourglass symbol. 

#### Warning and Safety notes:

For safety and environmental information cf. safety data sheets. These are available from Miele websites listed below.

Dispose of containers only when empty and sealed. Disposal of product residue: refer safety data sheet.

All information contained herein is current and correct to our knowledge. This does not however release users from obligations to perform their own trials and experiments. This cannot be deemed as a legal warranty of product characteristics or properties.

#### Packaging units and accessories:

Designation	Packaging unit	Mat. no
<i>ProCare Dent 10 A</i>	5 l (1.32 gal) canister	12 186 660
Canister key	For 5 l and 20 l canisters	11 851 600
DTD2 Drip tray	For 5 l and 10 l canisters	10 695 460

#### Contact information:

##### Germany

Miele & Cie. KG  
Carl-Miele-Straße 29,  
33332 Gütersloh  
Tel.: +49 (0) 8002244644  
vertrieb@miele-professional.de  
www.miele.de/professional

##### USA

Miele Inc.  
9 Independence Way  
Princeton, NJ 08540  
Tel.: 800-991-9380  
www.miele-professional.com



##### Germany

Chemische Fabrik  
Dr. Weigert GmbH & Co. KG  
Mühlhagen 85  
20539 Hamburg, Germany  
www.drweigert.de