

CERTIFICATE

Prof. Dr. Dirk Bockmühl
Campus Kleve

Telefon: +49 (0) 28 21 / 806 73-208

Telefon (Sekretariat): +49 (0) 28 21 / 806 73-213 oder -242

Fax: +49 (0) 28 21 / 806 73-260

E-Mail: dirk.bockmuehl@hochschule-rhein-waal.de

(stets angeben) Zeichen: DB

Datum: 27. November 2019

This is to certify that the PG 8063 AE dishwasher manufactured by
Miele & Cie KG, Carl-Miele-Strasse 29, 33332 Gütersloh, Germany

has exceeded the requirements for germ reduction by 5 log, as requested for chemical disinfectants, in the tests carried out by the Rhein-Waal University of Applied Sciences / Faculty of Life Sciences (test report HM020005101, dated 31.10.2019).

The average germ reduction achieved in the program "Sf Special" for reprocessing of Dräger and MSA respirator masks was 7.29 log, and 6.92 log levels for the program "Variochem" for reprocessing of Interspiro respirator masks.

Kleve, 27.November 2019



Prof. Dr. Dirk Bockmühl